


Product Specifications Sheet

Specification Sheet No	SSN-FT1357-008			  Remark A - Product length B - Product width C - PU length D - PU to edge
Product Name	Wound Care			
Product Description	Transparent Film Dressing (365 Film Dressings)			
Range	365 Healthcare			
Manufacturer	Shanghai ISO Medical Products Co., Ltd.			
Material	Polyurethane (PU) film with thickness Pressure-sensitive adhesive			
Sterilisation	EO Sterile			
Issue Date	13 th January 2026	Version No	R03	

Intended use	A single use transparent film dressing intended to be used as a primary dressing for superficial wounds or as a secondary fixation.
--------------	---

Packaging								
Product code	Colour	A (mm)	B (mm)	C (mm)	D (mm)	Items Per Pack	Qty- Inner	Qty – Shipper Carton
36590006	Transparent	69 ±2	40 ±2	50 ±2	10 ±2	1	100	2000
36590007		100 ±5	60 ±3	71 ±3	14.5 ±3		100	2000
36590008		152 ±3	95 ±3	122 ±2	16 ±2		50	800
36590009		242 ±3	100 ±3	250 ±3	15 ±2		30	300
36590010		280 ±3	100 ±3	250 ±3	15 ±2		30	300
36590051		242 ±3	147 ±3	200 ±3	21 ±2		25	250
36590052		152 ±3	95 ±3	140 ±2	6 ±2		50	800
36590053		215 ±3	200 ±3	200 ±3	15 ±2		200	400
36590054		320 ±5	205 ±4	305 ±5	/		10	100

Quality Certification	ISO 13485:2016		
Device Classification	Class IIa	Rule	Rule 4
Conformity Route	Medical Device Directive 93/42/EEC		
Storage conditions	Store in a cool dry place away from heat and direct sunlight		
Shelf life	5 years if stored as recommended		
Contraindications/Warnings	None		
Markings	Ref, Product name, Size, Image, Quantity, Barcodes, Lot, Manufacturing date, Expiry date, CE Mark, Single use, Latex free, EO Sterile, Medical device, Do not use if package is damaged, consult instruction for use, keep dry, keep away from direct sunlight		
Additional information	It is showerproof, enabling patients to shower without disturbing their dressings. It is breathable allowing effective oxygen and moisture vapour exchange (MVTR 1176 g/m2/24 hours) to optimise the wound healing environment. Latex free and sterile. Do not use if inner packaging is broken.		

Medical Device and Reporting System: Incidents caused by this device must be reported immediately to our offices.

The information provided within this Specification Sheet is correct as of last revision but it is subject to change without notice.

365 Healthcare	Mediq House
T: +44 (0) 1332 413 700	Arundel Avenue, Castle Donington
W: https://mediq.co.uk/	Derby, DE74 2HJ

365 Healthcare is a trading division of Mediq Healthcare UK Ltd. Registered in England No. 62537

Document Reference	Document Name	Reviewed By	Approved By	Revision Number	Issue Date
DMS 04-004	Product Specification Sheet	Jack Wong	Racheal Begg	R02	03 rd April 2024

Amendment Record

This procedure is reviewed to ensure its continuing relevance to the system and process that it describes. A record of contextual additions or omissions is given below. This revision history lists all changes to this document, the date of each revision, and the reason for each revision.

Revision Number	Revision Date	Reason for Revision	Updated by
R01a	17/07/2024	Update to Mediq Healthcare UK brand replacing DSN681	Elizabeth Ogunkolade
R02a	12/02/2025	Update to revision 2	Elizabeth Ogunkolade
R03	13/01/2026	Update the shipper carton qty	Jack Wong

This document is uncontrolled when printed. The master document is controlled electronically, and document users are responsible for ensuring printed copies are valid prior to use to ensure it represents current practice within Mediq Healthcare UK Ltd.