

Product Procedure Specifications Sheet

Specification Sheet No	SSN-FT1390-005		
Product Name	Umbilical Cord Clamp Pack		
Product Description	36519559 (Rev3)		
Range	365 Healthcare		
Manufacturer	Wujiang Kangxin Surgical Dressings Co., Ltd.		
Packaging	100pcs/case		
Sterilisation	EO Sterile		
Issue Date	17th July 2024	Version No	R02



IMAGE
COMING SOON

Intended use	This range consists of general supplementary items which can be used on its own or in addition to other procedure packs for the required clinical procedure.
--------------	--

Pack content				
Item	Product code	Quantity	Dimensions	Material
Umbilical Cord Clamp	4802	1	5 cm	Polyformaldehyde
White non-woven swab 4ply	3103	1	10 x 10 cm	Viscose & terylene

Quality Certification	ISO 13485:2016		
Device Classification	Class Is	Rule	4
Conformity Route	Medical Devices Regulation (EU) 2017/745 Medical Devices Regulations 2002 (SI 2002 No 618, as amended) (UK MDR 2002)		
Storage conditions	Store in a cool dry place away from heat and direct sunlight.		
Shelf life	5 years if stored under recommended storage conditions		
Contraindications/Warnings	None		
Markings	Product Name, REF, 365 Logo and Address, Lot Number, Manufacturing Date, Expiry Date Medical Device, Single Use, EO Sterile, Single sterile barrier, do not use if package damaged, Keep dry and away from sunlight, CE, ECREP, Manufacturer Name and Address, Distributor Name and Address		
Additional information	Products are latex and phthalate free. Products contain no hazardous substances. Products contain no medical substances.		
Medical Device and Reporting System: Incidents caused by this device must be reported immediately to our offices.			

The information provided within this Specification Sheet is correct as of last revision but it is subject to change without notice.

365 Healthcare	Mediq House
T: +44 (0) 1332 413 700	Arundel Avenue, Castle Donington
W: https://mediq.co.uk/	Derby, DE74 2HJ

365 Healthcare is a trading division of Mediq Healthcare UK Ltd
Registered in England No. 62537

Document Reference	Document Name	Reviewed By	Approved By	Revision Number	Issue Date
DMS 04-004-03	Procedure Pack Specification Sheet	Jack Wong	Racheal Begg	R02	03 rd April 2024

Amendment Record

This procedure is reviewed to ensure its continuing relevance to the system and process that it describes. A record of contextual additions or omissions is given below. This revision history lists all changes to this document, the date of each revision, and the reason for each revision.

Revision Number	Revision Date	Reason for Revision	Updated by
R01a	17/07/2024	Update to Mediq Healthcare UK brand replacing DSN464	Elizabeth Ogunkolade

This document is uncontrolled when printed. The master document is controlled electronically, and document users are responsible for ensuring printed copies are valid prior to use to ensure it represents current practice within Mediq Healthcare UK Ltd.