# Product Specifications Sheet

| | | | |
|---|---|---|---|
| **Specification Sheet No** | SSN-JFRU06-001 | | |
| **Product Name** | UHS Suction Tube | | |
| **Product Description** | UHS Suction Tube | | |
| **Range** | Universal | | |
| **Manufacturer** | Romsons International | | |
| **Material** | Polyvinyl chloride (PVC) | | |
| **Sterilisation** | ETO sterilization | | |
| **Issue Date** | 21st May 2024 | **Version No** | R02 |



| | |
|---|---|
| **Intended use** | Used for connecting suction sources to suction waste collection systems, suction catheters, Yankauer suckers, suction probes and other suction devices. |

| **Packaging** | | | | |
|---|---|---|---|---|
| **Product code** | **Connector Colour** | **Tube Dimension (Internal diameter - length)** | **Connectors type** | **Qty – Shipper Carton** |
| UN44-6V-50 | Green | 7mm – 180cm | FFM Vacuum | 50 |
| UN44-8V-50 | Green | 7mm – 250cm | FFM Vacuum | 50 |

| | | | |
|---|---|---|---|
| **Quality Certification** | ISO 13485 | | |
| **Device Classification** | Class IIa | **Rule** | 2 |
| **Conformity Route** | MDD 93/42/EEC | | |
| **Storage conditions** | Store in a cool dry place away from heat and direct sunlight | | |
| **Shelf life** | 5 Years if stored as recommended above | | |
| **Contraindications/Warnings** | Peel Open the sterile pack with sterile gloved hands Only. | | |
| **Markings** | Product name, Universal Logo, Code Ref, lot number, expiry date, CE mark, irradiation, single use and latex free symbols. | | |
| **Additional information** | Non conductive<br>The products are DEHP free, latex free and single use.<br>Do not use if the packaging is wet or damaged. | | |
| *Medical Device and Reporting System: Incidents caused by this device must be reported immediately to our offices.* | | | |

The information provided within this Specification Sheet is correct as of last revision but it is subject to change without notice.


| | |
|---|---|
| 365 Healthcare | Mediq House |
| **T**: +44 (0) 1332 413 700 | Arundel Avenue, Castle Donington |
| **W**: https://mediq.co.uk/ | Derby, DE74 2HJ |

365 Healthcare is a trading division of Mediq Healthcare UK Ltd
Registered in England No. 62537

## Amendment Record

This procedure is reviewed to ensure its continuing relevance to the system and process that is describes. A record of contextual additions or omissions is given below. This revision history lists all changes to this document, the date of each revision, and the reason for each revision.

| Revision Number | Revision Date | Reason for Revision | Updated by |
|---|---|---|---|
| R01 | 17/11/2023 | Update to Mediq Healthcare UK brand Template | Jack Wong |
| R02 | 21/05/2024 | Update Mediq Healthcare UK address | Jack Wong |
| | | | |
| | | | |

This document is uncontrolled when printed. The master document is controlled electronically, and document users are responsible for ensuring printed copies are valid prior to use to ensure it represents current practice within Mediq Healthcare UK Ltd.