

Product Specifications Sheet

Specification Sheet No	SSN-JFRU06-006			
Product Name	Suction			
Product Description	UHS Suction Tubing 3mtr X 6mm FFM			
Range	Universal			
Manufacturer	Romsons International			
Material	Polyvinyl Chloride and Polypropylene			
Sterilisation	EO Sterile			
Issue Date	21st May 2024	Version No	R02	

Intended use	Intended for connecting suction sources to suction waste collection systems, suction catheters, Yankauer suckers, suction probes, and other suction devices.
--------------	--

Packaging				
Product Code	Length (Approx.)	Internal Diameter (Approx.)	Items Per Pack	Qty – Shipper Carton
UN30026FFM-50	300 cm	6 mm	1	50

Quality Certification	ISO 13485:2016		
Device Classification	Class IIa	Rule	2
Conformity Route	Medical Devices Directive (93/42/EEC)		
Storage conditions	Store in a cool dry place away from heat and direct sunlight		
Shelf life	5 Years from manufacturing date if stored as recommended		
Contraindications/Warnings	To be used by a qualified / a well-trained medical person.		
Markings	Universal Logo, Ref, Product name, Lot, Quantity, Barcodes, Manufacture date, Expiry date, CE Mark, Single use, Do not re-sterilise, Latex free, EO -sterile, DEHP free, Medical device, Do not use if package is damaged, keep dry, keep away from sunlight, Manufacturer Name & Address, Distributor contact, Non-pyrogenic		
Additional information	Product is packed in a peelable blister pack per unit. Female/Female and a Male Adaptor. Product is Latex-free		
Medical Device and Reporting System: Incidents caused by this device must be reported immediately to our offices.			

The information provided within this Specification Sheet is correct as of last revision but it is subject to change without notice.

365 Healthcare	Mediq House
T: +44 (0) 1332 413 700	Arundel Avenue, Castle Donington
W: https://mediq.co.uk/	Derby, DE74 2HJ

365 Healthcare is a trading division of Mediq Healthcare UK Ltd
Registered in England No. 62537

Document Reference	Document Name	Reviewed By	Approved By	Revision Number	Issue Date
DMS 04-004	Product Specification Sheet	Jack Wong	Racheal Begg	R02	03 rd April 2024

Amendment Record

This procedure is reviewed to ensure its continuing relevance to the system and process that it describes. A record of contextual additions or omissions is given below. This revision history lists all changes to this document, the date of each revision, and the reason for each revision.

Revision Number	Revision Date	Reason for Revision	Updated by
R01	15/02/2024	New SSN Mediq Healthcare UK brand Template	Elizabeth Ogunkolade
R02	21/05/2024	Update Mediq Healthcare UK address	Jack Wong

This document is uncontrolled when printed. The master document is controlled electronically, and document users are responsible for ensuring printed copies are valid prior to use to ensure it represents current practice within Mediq Healthcare UK Ltd.